

VIA Thaw™ automated dry thawers

CELL THERAPY SYSTEMS

VIA Thaw™ dry thawers provide automated, controlled and consistent thawing of cellular products in cryobags. These instruments use a dry heat technology and multiple sensors to determine when thawing is complete, removing the uncertainties associated with non-controlled systems and manual operations. VIA Thaw™ systems can also be integrated with Chronicle™ automation software¹, allowing user-customizable thawing profiles to be synchronized among multiple units and for the generation of complete electronic batch records.

VIA Thaw™ units offer the following benefits:

- **Consistent, automated thawing:** Automating thawing to an exact endpoint can maximize cell viability and minimize process variation and mishandling. Users can optimize thawing profiles and, through integration with Chronicle™ software, instantly share them to all linked units.
- **Multiple sizes:** Dry thawing technology applies the same process to multiple bag brands and volumes, from 10 to 275 mL.
- **Scalable thawing:** By connecting VIA Thaw™ instruments through Chronicle™ software, users can synchronize profiles to new VIA Thaw™ units when more thawing capacity is needed. Scale effectively, without changing processes.

The VIA Thaw™ L1000 system is intended for research or manufacturing use only and is not designed for clinical procedures or diagnostic purposes. VIA Thaw™ L1000 units should only be used by specifically trained scientists or technicians in a laboratory or manufacturing setting.

Instrument specifications and components

VIA Thaw™ instruments accommodate multiple cryobag sizes and fill volumes. Single cryobags are positioned in the tray, and the flexible bed molds to the shape of the frozen bag for efficient heat transfer (Fig 2).




Fig 1. The VIA Thaw™ L1000 instrument.



Fig 2. VIA Thaw™ L1000 with thawing tray removed to show the flexible bed.


Once thawing starts, agitators (on the top plate) gently agitate the cryobag surface. Infrared temperature sensors monitor the temperature trend, triggering thawing process steps and alarms as needed. Indicator lights on the control panel change in correspondence to advancing process steps.

Users can customize thawing profile parameters for enhanced process control, such as bag fill volume, thawing time, initial temperature alarm level, and accessibility. Alarms alert operators when defined temperature setpoints are exceeded or when thawing is complete. Alerts can be selectively enabled on VIA Thaw™ L1000 units.



| Part | Description |
|------|----------------------------|
| 1 | Thawing tray |
| 2 | Top heating plate |
| 3 | Barcode reader |
| 4 | Infrared (IR) sensors (×5) |
| 5 | Cryobag agitator (×4) |
| 6 | Grip |
| 7 | Lid |
| 8 | Bottom heating plate |
| 9 | Control panel |
| 10 | Drain plunger |

Fig 3. Illustration showing an overview of VIA Thaw™ instruments.



| Part | Description |
|------|-------------------|
| 1 | Application inlet |
| 2 | Fuse holder |
| 3 | Power switch |
| 4 | Network port |

Fig 4. Illustration showing the power and network connections present on VIA Thaw™ instruments.

Specification table

| | |
|--|--|
| Maximum cryobag dimensions (W × L × H) | 305 × 140 × 15 mm |
| Compatible cryobags manufacturers* | Charter Medical, OriGen, Miltenyi™ |
| Cryobag fill volume | 10 to 275 mL |
| Instrument dimensions (W × L × H) | 430 × 375 × 150 mm |
| Weight | 13.0 kg |
| Thaw loading temperatures | From -196°C to 0°C |
| Thawing profiles | <ul style="list-style-type: none"> • 8 pre-programmed profiles available and editable • 8 profiles downloadable from Chronicle™ software |
| Networking† | Wired LAN: IEEE 802.3 Wireless LAN: IEEE 802.11 b/g |
| Thawing temperature | 34 °C ± 1 °C |
| Integrated camera | Yes |
| Operating voltage | 100-240 VAC (50/60 Hz) |
| Maximum power consumption | 650 VA |
| Connects to Chronicle™ automation software | Yes |

* Contact your local Cytiva representative for detailed information about cryobag models and fill volumes.

† Wireless LAN is not available in all countries and is disabled by default on delivery. Wireless LAN can be activated in an approved region.

Ordering information

| Product | Product code |
|-----------------|--------------|
| VIA Thaw™ L1000 | 29403109 |
| Thawing tray | 29458851 |

Safety and compliance

The VIA Thaw™ L1000 instrument is manufactured and controlled under an ISO 9001 certified quality system. This unit is intended for research or manufacturing use only and is not cleared or approved for clinical procedures or diagnostic purposes. The VIA Thaw™ L1000 instrument has the following approvals: CE, cTUVus, FCC, IC, RCM, KC, EAC and UKCA.²

Manufacturer

Asymptote Ltd.,
 Sovereign House, Cambridge, CB24 9BZ
 United Kingdom
 +44(0) 3306 600 320

References

1. Chronicle™ automation software is a web-based, fit-for-purpose digital solution to view and store information about instruments, processed samples, and thawing profiles. Chronicle™ software is not part of VIA Thaw™ instruments, is not required for instrument functioning, and can be purchased separately.

2. CE = Compliance for European Union and European Economic Area; cTUVus = NRTL (Nationally Recognized Testing Laboratory) certification from TÜV Rheinland, demonstrating compliance for US and Canadian markets; FCC = Federal Communications Commission; IC = Industry Canada; RCM = Regulatory Compliance Mark for Australia and New Zealand; KC = Korea Certification; EAC = EurAsian Conformity; UKCA = UK Conformity Assessed.

[cytiva.com/celltherapy](https://www.cytiva.com/celltherapy)

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdings Corporation or an affiliate. Chronicle and VIA Thaw are trademarks of Global Life Sciences Solutions USA LLC or an affiliate.

Miltenyi is a trademark of Miltenyi Biotec. All other third-party trademarks are the property of their respective owners.

© 2021 Cytiva

For local office contact information, visit [cytiva.com/contact](https://www.cytiva.com/contact)

CY12085-21July21-DF

